



Series 2701 System SourceMeter® Instruments

Version B02 Installation Considerations

Keithley Instruments, Inc.
28775 Aurora Road
Cleveland, Ohio 44139-1891
1-888-KEITHLEY
www.keithley.com

Important Upgrade / Downgrade Installation Considerations:

- Upgrading from Axx version to B01 or B02 is possible, but offers no benefit.
- Do not downgrade B01 or higher into units with Axx version.
- Contact Keithley Application Group for the Firmware upgrade file.

B02 Release

Overview

The B02 release addressed an ethernet connection failure seen on some B01 units. The upgrade additionally requires a hardware modification in order to correct the symptom, so no benefit is gained by upgrading unless the hardware modification is also applied.

Instrument should be returned to Keithley for the upgrade and modification.

B01 Release

Overview

The B01 release was necessary to support hardware changes in the Model 2701, and is backward compatible with older hardware. No enhancements or bug fixes were made in B01.

Do not downgrade from B01 or higher as the ethernet connection may not work.